



STANDARDS
MALAYSIA

SKIM AKREDITASI MAKMAL MALAYSIA (SAMM)
LABORATORY ACCREDITATION SCHEME OF MALAYSIA

**SC 1.5 – SPECIFIC CRITERIA FOR ACCREDITATION OF
MECHANICAL TESTING AND NON-DESTRUCTIVE
TESTING (NDT)**

Issue 2, 30 April 2020

(Supplementary to MS ISO/IEC 17025)



MS ISO/IEC 17025

JABATAN STANDARD MALAYSIA
Department of Standards Malaysia

TABLE OF CONTENTS

	Page
	1
1	1
2	1
3	1
4	1
5	1
6	2
6.1	2
6.2	2
6.3	3
6.4	4
6.5	5
6.6	5
7	5
7.1	5
7.2	5
7.3	6
7.4	6
7.5	6
7.6	6
7.7	6
7.8	6
7.9	6
7.10	6
7.11	6
8	6
Appendix 1	8
Appendix 2	15
Bibliography	24
Acknowledgements	

Introduction

The SC 1.5 document sets out the specific requirements for a mechanical testing and/or Non-Destructive Test (NDT) laboratory need to comply with.

This document shall be read in conjunction with MS ISO/IEC 17025, *Skim Akreditasi Makmal Malaysia* (SAMM) policies and other relevant requirements published by Department of Standards Malaysia (Standards Malaysia).

The clause numbers in this document correspond to those of MS ISO/ IEC 17025 but since not all clauses require additional requirements, the numbering may not be continuous.

1 Scope

Standards Malaysia accreditation does not constitute a blanket approval of all laboratory's activities. Therefore, it is necessary to identify those activities for which accreditation are granted. The classes of test provide the framework within which the scope of accreditation is expressed.

These classes and subclasses do not constitute any restriction on the work that a laboratory can perform, but provide a convenient means of expressing a laboratory's recognised capability.

Classes of test appropriate to mechanical testing and NDT laboratories are listed in **Appendix 1**. These classes are an arbitrary subdivision of the potential range of activities involved in mechanical testing and NDT laboratories on the basis of the types of samples being tested, the scientific disciplines involved, and the test methods employed.

2 Normative reference

MS ISO/IEC 17025 - General requirements for the competence of testing and calibration laboratories.

The undated references indicate the latest edition of the referenced documents, including any amendments.

3 Terms and definitions

None.

4 General requirements

Same as in MS ISO/IEC 17025.

5 Structural requirements

Same as in MS ISO/IEC 17025.

6 Resource requirements

6.1 General

Same as in MS ISO/IEC 17025.

6.2 Personnel

6.2.1 Mechanical and/or NDT testings shall be performed by competent personnel or supervised by authorised personnel.

6.2.2 On-going competence should be monitored on regular basis. Records on competency shall be available. Where a method or technique is not in regular use, verification of personnel performance before testing is necessary.

6.2.3 Particular attention shall be given to the following aspects:

- (a) There shall be clearly defined and recognisable lines of authority and responsibility within the organisation, with each personnel being aware of both the extent and the limitations of their own responsibility.
- (b) Laboratory personnel shall only be allocated duties that commensurate with their knowledge and experience. They shall be provided with the direction or supervision needed for effective performance of their duties. Authorisation and up to date competency records (training records) shall be available for all personnel carrying out testing work.
- (c) Competency of personnel is assessed by peer review during practical demonstrations. Some areas of expertise, where tests involve technical judgement e.g. NDT, shall meet standard or regulatory requirements as prerequisites.

6.2.4 Approved Signatory

Same as requirements in clauses 4 and 5 of SAMM Policy (SP6) - Requirements for SAMM Approved Signatory. In addition,

- (a) The laboratory shall have at least one laboratory personnel who is qualified as approved signatory in the testing being undertaken. Requirements for Approved Signatories are detailed in SAMM Policy 6.
- (b) The signatories for NDT shall be qualified, have 2 years of relevant supervisory experience and certified to **minimum**:
 - i) NDT certification Level II of *Sijil Kemahiran Malaysia* by **Jabatan Pembangunan Kemahiran** (Department of Skills Development) or equivalent NDT certification schemes by any qualification bodies certified to ISO 9712, such as Personnel Certification in Non-Destructive Testing (PCN) by British Institute for Non-Destructive Testing (BINDT) or ASNT Central Certification Program (ACCP) by American Society for Non-destructive Testing (ASNT) or Australian

Institute for Non-Destructive Testing (AINDT), etc.

- ii) In addition to i) above, personnel performing radiographic testing shall obtain approval from the Atomic Energy Licensing Board (AELB).

Note:

The requirements of AELB may be referred in <http://www.aelb.gov.my/malay/dokumen/notis-pemberitahuan/Bil032013.pdf>. The latest edition of the referenced document, including any amendments applies.

6.3 Facilities and environmental conditions

Same as requirements in MS ISO/IEC 17025.

6.3.1 When highly precise measurements are to be made, the following factors may assume greater importance:

- (a) Isolation from sources of mechanical vibration and shock likely to have a detrimental effect on sensitive instruments (e.g. high accuracy balances).
- (b) Adequate ventilation when fumes are created by the tests such as in bitumen testing.
- (c) Temperature and humidity control of the laboratory as specified in the relevant test method (e.g. paper testing).
- (d) Protection from excessive levels of dirt and dust.
- (e) Suitable equipment and areas for the preparation of test specimens such as in tensile testing and metallography.
- (f) Isolation from stray electric and magnetic fields, particularly for thermocouples, strain gauges and other sensitive low output devices.
- (g) Electromagnetic interference between items of test equipment and computers.

6.3.2 Safety

Some types of tests have very specific safety requirements, which shall be met, e.g. radiographic, and these may be subject to regulatory requirements.

Other tests will have less specific but otherwise significant safety concerns, e.g. compression tests on concrete. It is expected that accredited laboratories will have considered, and provide appropriate safety procedures to cover items such as:

- (a) Noise - from equipment such as mechanical sieve shakers and compaction hammers.
- (b) Ventilation - adequate air flows in controlled environments - protection from corrosive or toxic fumes.

- (c) Personal Protection - safety clothing, etc.
- (d) Physical Protection - safety screens on equipment such as compression testers.

Note:

Occupational Safety and Health Act 1994 (Act 514) and Factories and Machinery Act 1967-revised 1974 (Act 139) places specific legal obligations on all employers, including laboratories. Safety is outside the scope of accreditation and will not be assessed during an on-site laboratory accreditation assessment. If, in the opinion of the assessment team, a safety issue is observed during an assessment, it will be reported to the laboratory, as required by the Acts. The reporting of a safety issue will not indicate that a comprehensive safety assessment has been carried out.

6.4 Equipment

Same as in MS ISO/IEC 17025.

- 6.4.1 Guidelines on calibration requirements and recalibration intervals for equipment are detailed in **Appendix 2**. The guidelines set out **maximum** periods of use before equipment must be recalibrated. Further details are in ILAC-G24 Document.
- 6.4.2 Reduced or extended calibration intervals may be accepted based on factors such as history of stability, accuracy required and ability of personnel to perform regular checks. It is the responsibility of the laboratory to provide clear evidence that its calibration system, and any changes to an existing system, will ensure that confidence in the equipment can be maintained.
- 6.4.3 Equipment that is sensitive to movement, such as force, impact, hardness testing machines, heat enclosures and balances, will generally require full recalibration if they are moved.
- 6.4.4 A laboratory, which uses a computerised testing system, shall comply with the following criteria:
 - (a) The optimum calibration procedure for physical testing systems will depend upon the accessibility of individual components of the system, especially their input or output signals.

If a testing instrument **cannot be** isolated from the data processing system, the system as a whole shall be calibrated either statically or dynamically. Calibrating the complete system is the preferred alternative.

If the testing instrument **can be** isolated from the data processing system, each component of the system can be calibrated or verified separately. The testing instrument can be calibrated (again, statically, or dynamically) in the conventional manner and a separate verification of the data processing system, including any interfacing systems, can be undertaken.
 - (b) The computer programme should be comprehensive in its coverage of the testing process and should have been checked at points covering the

whole range of likely inputs and outputs.

- (c) The programme should allow the operator to detect errors in data input and to monitor the progress of the test.
- (d) The system should be capable of being checked for error-free operation with respect to data capture, data processing, and freedom from sources of external interference. Where appropriate, manually checked data sets (or artefacts) should be available for regular system checks.

6.5 Metrological traceability

Same as in MS ISO/IEC 17025 and SAMM Policy (SP2) - Policy on the Traceability of Measurement Results.

6.6 Externally provided products and services

Same as in MS ISO/IEC 17025.

7 Process requirements

7.1 Review of requests, tenders and contracts

Same as in MS ISO/IEC 17025.

7.2 Selection and verification of methods

- (a) Accreditation is normally granted only for internationally or nationally accepted standard test methods or non-standard methods (in-house methods) that have been appropriately verified/validated, and which are performed regularly. Refer to **Table 1**. The extent of a laboratory's scope of accreditation will therefore vary with the range of work performed, the scope and complexity of the tests involved, the competence and organisation of laboratory personnel and the level of technology available in the laboratory.
- (b) In-house methods could include but are not restricted to:
 - i) methods developed in the laboratory;
 - ii) methods developed by customer/manufacturer;
 - iii) methods developed for an industry group;
 - iv) modified standard test methods; and
 - v) method published in the scientific literature with/without any performance data.
- (c) Validation of test methods shall involve, as appropriate, the use of certified reference materials, participation in inter-laboratory comparison/proficiency test programmes, comparison with standard test methods, determination of method precision, limits of detection, uncertainties of

measurement, etc.

(d) Standard test methods should be used whenever possible in order to ensure comparability of test results among laboratories.

(e) For NDT, the procedures shall be approved and authorised by NDT level III personnel. The NDT instructions shall be approved by either a Level II or a Level III personnel.

7.3 Sampling

Same as in MS ISO/IEC 17025.

7.4 Handling of test and calibration items

Same as in MS ISO/IEC 17025.

7.5 Technical Records

Same as in MS ISO/IEC 17025.

7.6 Evaluation of measurement uncertainty

Same as in MS ISO/IEC 17025 and SAMM Policy (SP5) - Measurement Uncertainty Requirements for SAMM Testing Laboratories.

Evaluation of measurement uncertainty is not required where test results are qualitative (i.e. nonnumeric such as pass/fail or fracture/no fracture), unless it is required by a testing standard or customer's specification.

7.7 Ensuring the validity of results

Same as in MS ISO/IEC 17025 and SAMM Policy (SP4) - Policy for participation in proficiency testing activities.

7.8 Reporting the results

Same as in MS ISO/IEC 17025.

7.9 Complaints

Same as in MS ISO/IEC 17025.

7.10 Nonconforming work

Same as in MS ISO/IEC 17025.

7.11 Control of data and information management

Same as in MS ISO/IEC 17025.

8 Management system requirements

Same as in MS ISO/IEC 17025.

Table 1: Validation or Verification requirements based on methods selected

Test Method Description		Validation or Verification requirements	Method Reference No/ID (Example)
Standard published method		Confirmation of published performance characteristics	<Method>, <year/edition>, <section no.> e.g.: MS 1: 1996 Clause 12.2 ASTM A370-03a
In-house test method			
Method developed by laboratory		Full validation	<In-house method>, <ref. no.> e.g.: In-house method: 1234-A
Method developed by customer/manufacturer		Full validation	<In-house method>, <ref. no.><based on xxxx > In House Method e.g.: In-House Method MGT/001/ES-X 60210 Based on MS 30: Part 5: 1995, Section 2
Method developed for an industry group	Absence of performance characteristics	Full validation	<Method>, <year/edition> e.g.: ASME VIII UCS (56)
	Confirmation of published performance characteristics		<Method>, <year/edition>, <section no.> e.g.: API 2H-1993 Annex S-4 ASME B30.20: 2006 Section 20-1.3.8
Modified standard methods		Full validation	<In-house method>, <ref. no.>, <based on std. method>, <technique (sample preparation & detection, where applicable)> e.g.: In-house method QMCL/014/2007 based on MS 522: Part 2:2005 Clause 3
Method published in the scientific literature with any performance data		Confirmation of published performance characteristics	<In-house method>, <ref. no.>, <based on xxxx > e.g.:
Method published in the scientific literature without any performance data		Full validation	In house method ABC, based on Practical Guidebook for Radioisotope-based Technology in Industry, IAEA/RCA RAS/8/078 March 1999.

Appendix 1

Classes of Testing: Mechanical and NDT

A. Mechanical Testing

Mechanical and physical testings of material/ products that include metallurgical tests to determine the elemental analysis and microstructures.

Note: * All tests referring to relevant products standard

1. Metals and metal products

- Bend and re-bend
- Brinell hardness
- Charpy impact
- Compression, transverse and shear
- Drop-weight
- Fracture toughness
- Micro hardness
- Rebound hardness
- Rockwell hardness
- Stress-rupture
- Superficial Rockwell hardness
- Vickers hardness
- Tensile
- Others

2. Welds and welded test specimens

- Bend
- Corrosion
- Cracking
- Drop-weight
- Fillet-break
- Fracture toughness
- Hardness
- Impact
- Macroscopic examinations
- Nick-break
- Shear
- Tension
- Others

3. Lifting gear, chain, wire rope and fittings

- Breaking Load
- Proof load
- Tension
- Others

4. Fibre rope and cordage

- Tension
- Others

5. Springs and energy absorbing devices

- Compression
- Tension
- Torsion
- Others

6. Threaded fasteners

- Dimension
- Drive
- Proof load
- Stripping
- Tensile
- Tension-torque
- Torsion
- Others

7. Ceramic products

- Abrasion
- Crazing
- Dimensional
- Flushing
- Glazing
- Loading
- Marking
- Staining
- Tolerance
- Warpage
- Others

8. Concrete (fresh and hardened)

- Abrasion resistance
- Air content
- Cement content
- Creep
- Compression
- Density
- Drying shrinkage
- Flexural strength
- Flow table
- Initial surface absorption (ISAT)
- Modulus of elasticity
- Rapid chloride permeability (RCPT)
- Sampling
- Setting time
- Slump
- Splitting tensile
- Standard consistence
- Water absorption
- Water permeability
- Others

9. Cement / concrete based products

- Abrasion
- Breaking
- Compressive strength
- Dimension
- Dynamic
- Fire propagation / resistance
- Flexural
- Load
- Sampling
- Shrinkage
- Splitting
- Water absorption
- Others

10. Refractories

- Cold crushing strength
- Density and porosity
- Durability
- Hydrogen diffusivity
- Modulus of elasticity
- Modulus of rupture
- Particle Size Determination (PSD)
- Others

11. Rocks

Compressive strength
Elastic moduli
Petrographic examination
Point load strength
Strength
Others

12. Cements and pozzolanic materials

Air content
Compressive strength
Fineness
Flexural
Sampling
Setting time
Soundness
Standard consistence
Water absorption
Others

13. Bituminous materials and Bituminous pavement (solid and liquid)

Abrasion
Bitumen extraction
Bitumen short term aging
Brittleness
Ductility
Elastic recovery
Flash and fire point
Float
Fraass breaking point
Long term aging (pressure aging vessel)
Loss on heating
Marshall stability
Penetration
Polish stone value
Rheology
Sampling
Softening point
Specific gravity
Thickness
Viscosity
Water content
Others

14. Soils

California Bearing Ratio (CBR)
Classification
Compaction (Proctor)
Consolidation
Density
Field density (FDT)
Linear shrinkage
Liquid limit
Moisture content
Outdoor weathering reactivity
Plastic limit
Sampling
Specific gravity
Strength
Others

15. Timber and timber products

Bending strength
Block shear
Compression strength
Delamination
Density
Flexural strength
Modulus of elasticity
Moisture content
Reaction to fire
Rolling shear strength and stiffness
Shear strength
Specific gravity
Stress grading timber
Tensile strength
Torsion
Others

16. Building boards and plywood

Adhesion of plies
Bonding
Density and moisture content
Fire resistance
Flexural strength
Joint strength
Load
Shear strength
Tensile strength
Water resistance
Others

17. Glass and glass products

Acid and alkaline
Annealing point and strain point
Boil
Dimension and shape
Drop ball
Emissivity
Fragmentation
Opacity
Pendulum impact
Softening point
Sound
Static puncture
Surface compression
Visible light/solar transmittance
Weathering
Others

18. Clays and clay products

Abrasion
Breaking strength
Performance
Chemical and staining
Modulus of Rupture
Compression
Crushing
Dimension
Flexural
Porosity and shrinkage
Water absorption
Water leakage
Others

19. Aggregates

Aggregate crushing value (ACV)
Aggregate impact value (AIV)
Angularity number
Bulk density and water absorption
Sieve analysis
Chloride content
Clay, silt and dust content
Degradation tests
Elongation
Fine particle size distribution
Flakiness index
Friable particle
Light weight particle
Los Angeles value
Methylene blue
Moisture content
Organic impurities
Petrographic examination
Polished stone value
Potential alkali reactivity by mortar bar
Sampling
Sand equivalent
Shell content
Soundness
Sulphur content
Ten percent fines value (TFV)
Wet/dry strength ratio
Others

20. Pulpwood, pulp, paper, paperboard and products

Adhesives
Burst
Compression
Liquid absorption
Mechanical properties
Optical properties
Permeability
Sampling
Surface properties
Tear
Tension
Others

21. Rubber and related products

Abrasion
Ageing and environmental
Belting
Brittleness
Compression
Curing characteristics with oscillating disc
Rheometer
Density and specific gravity
Elastomeric bearings
Electrical resistivity
Flammability
Flexing
Hardness
Low temperature
Mooney Accelerated Storage Hardening Test (MASHT)
Mooney viscosity
Ozone resistance
Plasticity retention index (PRI)
Sampling
Shear
Swelling in liquids
Tear
Tension
Tension set
Viscosity
Vulcanisation characteristics
Wallace Accelerated Storage Hardening Test (WASHT)
Others

22. Gypsum and gypsum products

Adhesion
Compressive strength
Core cohesion at high temperature
Dimension
Flexural strength
Flow table
Hardness
Setting time
Shear strength
Water absorption
Others

23. Textiles and related products

Colour fastness
Flammability
Sampling
Tension
Tear burst
Wear
Others

24. Tyres

Bead unseating resistance
Dimensional
Endurance
High speed performance
Load/speed
Plunger energy
Strength
Treadwear indicators
Others

25. Automotive parts

Accelerated exposure
Adhesion of coating
Cleanliness
Damp heat, cyclic
Damp heat, steady state
Dry heat
Heat aging
Heat resistance
Humidity resistance
Low temperature resistance
Random vibration
Rapid change of temperature
Resonance frequency detection
Salt Spray
Scoring and condensation
Solvent resistance friction
Solvent resistance immersion
Temperature cycle
Vibration Endurance
Vibration Function (sinusoidal)
Others

26. Seat belts and similar devices

Abrasion
Adjusting force
Breaking strength
Corrosion resistance
Durability
Dust resistance
Dynamic
Exposure to water
Microslip
Releasing force
Retracing force
Strength
Temperature conditioning
Tilt lock
Vehicle sensitivity
Webbing sensitivity
Others

27. Personal Protective Equipment (PPE)

a. Safety Footwear & Occupational Footwear

Abrasion resistance
Behaviour of toecaps
Breaking strength of shoelace
Compression resistance
Construction
Flexing resistance
Heat insulation
Height of the upper
Hydrolysis
Impact resistance
Insole thickness material
Interlayer bond strength outsole
Internal length of toecaps
Leakproofness
Outsole thickness
Penetration resistance
Seat region (design B, C, D, E)
Sole adhesion
Specific ergonomic features
Tear strength
Tensile properties
Thickness
Upper flexing resistance-Bally flex
Upper/outsole bond strength
Water absorption/desorption
Water vapour permeability and coefficient

b. Protective Helmet for Motorcyclist

Chin Strap Micro-slip
Detaching
Dynamic
Impact absorption
Impact Energy Attenuation
Penetration
Projection and surface friction
Resistance to abrasion
Retention
Rigidity

c. Protective Visor for Motorcyclist

Light diffusion
Luminous transmittance
Mechanical characteristics
Mist retardant
Optical quality and scratch resistance
Recognition of signal light
Refractive powers
Spectral transmittance

28. Packages and containers

Aperture and Closure
Compression
Dynamic
Fill Line Indicator
Free fall drop
Hydraulic pressure
Internal pressure (hydraulic)
Leakproofness
Mechanical – shock test
Penetration
Random vibration
Righting
Stability
Stacking
Strength of handle
Tear
Top Lift
Topple
Others

29. Environmental Tests

Conditioning
Erosion
Noise
Vibration
Others

30. Plastics and related products

Ageing and environmental
Bend
Burst
Compression
Elongation at break
Flammability
Flow properties
Hardness
Heat deflection temperature
Heat distortion
Hydrostatic internal pressure
Impact strength
Longitudinal/heat reversion
Low temperature
Melt mass-flow rate (MFR)
Sampling
Shear
Specific gravity
Tear
Tensile strength and Yield strength
Wear and abrasion
Others

31. Leather and leather products

32. Gasket, seals and packing

Fire
Gasket material
High pressure /temperature
Others

33. Adhesive and sealers

Adhesion of the coating (hot water soak test)
Cure
Peel strength
Others

34. Adhesive tapes

Peel adhesion strength
Adhesive bond strength
Others

35. Pipes and pipelines, hoses, valves and fittings

Acoustic
Adhesion
Burst
Cryogenic
Dimension
Endurance
Fire
Flow rate
Fugitive emission
Head loss
Holiday detector
Hydrostatic pressure
Leak tightness
Pneumatic pressure
Others

36. Mechanical tests on assemblies

Shear resistance of framed walls
Static test for lumber
Strength test of panels for building construction
Structural performance of exterior windows / doors / curtain walls, etc.
Wind load resistance
Others

37. Plumbing and drainage fittings

Internal pressure (hydrostatic pressure)
Longitudinal reversion
Noise test
Pendulum impact strength
Thermal cycling
Others

38. Furniture

Dimension
Durability
Impact
Safety
Stability
Strength
Others

Metallurgical

35. Microstructural tests on metallic & non-metallic alloys

Anodizing thickness
Case depth of surface defects
Depth of cladding
Depth of surface defects
Grain size
Hydrogen embrittlement
Macroscopic examination of steel
Macroscopic examination of wrought products
Non-metallic inclusion content
Proportion of size
Resistance to stress-corrosion cracking
Susceptibility of brass to dezincification
Others

39. Coatings

Abrasion
Acoustic
Adhesion strength
Alkaline resistance
Compressive strength
Corrosion resistance
Impact and scratch resistance
In-situ concrete lining
Salt spray
Surface area
Tensile strength
Thickness
Water absorption
Water pressure
Others

**40. Ferrous, non-Ferrous and metallic materials
Elemental Analysis (non-chemical method)**

Corrosion
Ferrite count
Metallography
Microstructure identification
Thermal Conductivity Detector Method (TCD Method)
Others

41. Mechanical equipment / toys and games/ sporting and recreational equipment / respiratory protective devices

Functional and performance
Endurance

42. Catalysts and catalyst carriers

Surface area
Pore size
Others

43. Other Tests

B. Non-Destructive Testing (NDT)

Examination of material, component and assembly to detect discontinuities without damaging the material, component or assembly.

Note: * All tests referring to relevant products standard

- 1. Metals and metal products**
 - Acoustic emission testing (AET)
 - Eddy current testing (ET)
 - Infrared thermographic testing (IRT)
 - Liquid penetrant testing (PT)
 - Magnetic particle testing (MT) (ferromagnetic only)
 - Radiographic testing (RT)
 - Ultrasonic testing (UT)
 - Visual testing (VT)
 - Others
- 2. Welds and welded test specimens**
 - Acoustic emission testing (AET)
 - Eddy current testing (ET)
 - Infrared thermographic testing (IRT)
 - Liquid penetrant testing (PT)
 - Magnetic particle testing (MT) (ferromagnetic only)
 - Radiographic testing (RT)
 - Ultrasonic testing (UT)
 - Visual testing (VT)
 - Others
- 3. Lifting gear, chain, wire rope and fittings**
 - Liquid penetrant testing (PT)
 - Magnetic flux leakage (MFL) (ferromagnetic only)
 - Magnetic particle testing (MT) (ferromagnetic only)
 - Radiographic testing (RT)
 - Ultrasonic testing (UT)
 - Visual testing (VT)
 - Others
- 4. Fibre rope and cordage**
 - Visual testing (VT)
 - Others
- 5. Springs and energy absorbing devices**
 - Liquid penetrant testing (PT)
 - Magnetic particle testing (MT) (ferromagnetic only)
 - Others
- 6. Threaded fasteners**
 - Liquid penetrant testing (PT)
 - Magnetic particle testing (MT) (ferromagnetic only)
 - Others
- 7. Ceramic products**
 - Liquid penetrant testing (PT)
 - Visual testing (VT)
 - Others
- 8. Concrete (fresh)**
 - Ultrasonic pulse velocity (UPV)
 - Others
- 9. Concrete (hardened)**
 - Carbonation test
 - Electromagnetic bar locator
 - Ground penetrating radar (GPR)
 - Infrared thermographic testing (IRT)
 - Radiographic testing (RT)
 - Rebound hammer
 - Ultrasonic pulse velocity (UPV)
 - Others
- 10. Cement / concrete based products**
 - Rebound hammer
 - Ultrasonic pulse velocity (UPV)
 - Others
- 11. Refractories**
 - Rebound hammer
 - Ultrasonic pulse velocity (UPV)
 - Others
- 12. Rocks**
 - Ultrasonic pulse velocity (UPV)
 - Discontinuity mapping
 - Terrain laser scanning (TLS)
 - Others
- 13. Cements and pozzolanic materials**
 - Ultrasonic pulse velocity (UPV)
 - Others
- 14. Bituminous materials and Bituminous pavement (solid)**
 - Ground penetrating radar (GPR)
 - Ultrasonic pulse velocity (UPV)
 - Others
- 15. Soils**
 - Nuclear density moisture gauge
 - Others
- 16. Timber and timber products**
 - Ultrasonic pulse velocity (UPV)
 - Others
- 17. Building boards and plywood**
 - Nuclear density moisture gauge
 - Others
- 18. Glass and glass products**
 - Ultrasonic pulse velocity (UPV)
 - Others
- 19. Rubber and related products**
 - Radiographic testing (RT)
 - Ultrasonic testing (UT)
 - Others
- 20. Tyres**
 - Radiographic testing (RT)
 - Ultrasonic testing (UT)
 - Others
- 21. Automotive parts**
 - Eddy current testing (ET)
 - Liquid penetrant testing (PT)
 - Magnetic particle testing (MT) (ferromagnetic only)
 - Radiographic testing (RT)
 - Ultrasonic testing (UT)
 - Others

22. Plastics and related products

Liquid penetrant testing (PT)
Ultrasonic testing (UT)
Others

23. Pipes and pipelines, hoses, valves and fittings

Liquid penetrant testing (PT)
Magnetic particle testing (MT) (ferromagnetic only)
Radiographic testing (RT)
Ultrasonic testing (UT)
Others

24. Mechanical assemblies

Eddy current testing (ET)
Liquid penetrant testing (PT)
Magnetic particle testing (MT) (ferromagnetic only)
Radiographic testing (RT)
Ultrasonic testing (UT)
Others

Metallurgical

25. Coatings

Eddy current testing (ET)
Others

26. Elemental Analysis (Non-chemical Method)

Eddy current testing (ET)
X-ray Fluorescent (XRF)
Others

27. Metal powders and sintered test

Liquid penetrant testing (PT)
Radiographic testing (RT)
Ultrasonic testing (UT)
Others

Mechanical Equipment

28. Cylinders and other pressure vessels

Acoustic emission testing (AET)
Infrared thermographic testing (IRT)
Liquid penetrant testing (PT)
Magnetic particle testing (MT) (ferromagnetic only)
Radiographic testing (RT)
Ultrasonic testing (UT)
Visual testing (VT)
Others

29. Fans and blowers

Vibration analysis
Others

30. Compressors

Vibration analysis
Others

31. Pumps

Vibration analysis
Others

32. Engines & generators

Vibration analysis
Others

33. Gas equipment & related products

Liquid penetrant testing (PT)
Magnetic particle testing (MT) (ferromagnetic only)
Radiographic testing (RT)
Ultrasonic testing (UT)
Others

Appendix 2

Equipment Calibration Intervals

Table 2 sets out the normal periods between successive calibrations for a number of reference standards and measuring instruments. It must be stressed that each period is generally considered to be the maximum appropriate in each case providing that the other criteria as specified below are met:

- a) The equipment is fit for purpose, and
- b) The laboratory has both the equipment capability and personnel expertise to perform adequate internal checks, and
- c) If any suspicion or indication of overloading or mishandling arises, the equipment is checked immediately and thereafter at frequent intervals until it can be shown that stability has not been impaired.

Where the above criteria cannot be met, appropriately shorter intervals may be necessary. It is possible to consider submissions for extension of calibration intervals based on factors such as history of stability, frequency of use, accuracy required and ability of personnel to perform regular checks. Application of the requirements of ISO 10012, Parts 1 and 2, need to be considered when seeking an extension of intervals. Where calibrations have been performed as above, adequate records of these measurements must be maintained.

Note: Checks or calibrations indicated * can be done internally by a laboratory providing they possess the necessary reference equipment, documented procedure and technical competence.

Table 2 a): Calibrations interval for reference standards and measuring instruments in mechanical testing

No.	Item of equipment	Calibration interval	Checking interval	General comments
1.	Accelerometers	One year		
2.	Anemometers	One year		
3.	Balances and Weighing Scales	Three years		By an accredited calibration authority. Twelve months service recommended
			*Each weighing	Zero check
			*One month	One-point check.
			*Six monthly	Repeatability check.
4.	Barometers	Three months (single point)		
5.	Dial Gauges	*Two years		
6.	Dies & Cutters			
7.	Extensometers			
	a) Level & mirror types	5 years		
	b) Micrometer screw type	5 years		
	c) Dial indicator type	2 years		
	d) Recording type (with electrical output)	2 years		

No.	Item of equipment	Calibration interval	Checking interval	General comments
8.	Force Testing Machines			
	Tension, Compression, & Universal	One year		Some Standards specify the recalibration period
	<u>TYPE 1 – Mechanical Force Measuring System</u>			
	a) Dead weight	Five years		
	b) Knife edge, lever and steelyard	Five years		
	c) Pendulum dynamometer	Two years		
	d) Chain testing and similar machines in frequent use	One year		
	<u>TYPE 2 – Hydraulic or Pneumatic Force Measuring Systems</u>			
	a) Mechanical system incorporating a pneumatic or hydraulic link, e.g. proportional cylinder	Two years		
	b) Bourdon Tube or diaphragm pressure gauge as force indicator	Six months		
	c) Type (b) fitted also with a master gauge which can be disconnected during normal testing	One year		Frequent checks by user of working gauge against master gauge
d) Bourdon tube or diaphragm gauge used only as a null detector for a mechanical system	Two years			
e) Bourdon tube with Measuring System	Two years			
<u>TYPE 3 – Electrical Force Measuring Systems</u>				
<u>Two years</u>				
9.	Gauge Blocks			
	a) Used as reference standards	Five years		
	b) Used as working equipment	Two years		More frequent as appropriate to usage.
10.	Hardness Testers for Metals			
	a) Brinell, Vickers and Rockwell machines	One year (partial) Three years (complete)	Daily check when in use	BS EN 10003 (Brinell), BS EN ISO 6507 (Vickers), BS EN 10109 (Rockwell)
	b) Portable Brinell microscopes	One year		with calibrated graticule
	c) Diamond indenters	*One year (inspection)		
11.	Hardness Testers for Rubber Plastics and Ebonite			
	a) Dead weight testers for rubber	Three years		BS 903 Methods N, A, L, M
	b) Dead weight testers for plastics	Three years		
	c) Meters (durometers) for rubber		Frequent checks by user on reference hardness blocks	
12.	Hydrometers	*Five years (one point)		BS 718

No.	Item of equipment	Calibration interval	Checking interval	General comments
13.	Hygrometers			
	a) Assman hygrometers and sling type	*Six months Five years (complete)		Compare thermometers at ambient with wick dry.
	b) Recorders accurate to $\pm 1\%$ RH	Two years		ASTM E77
	c) Other recorders including hair types	Weekly (with Assman hygrometer)		
	d) Digital instruments	One year		
14.	Impact Testing Machines (Pendulum type)			
	a) Charpy, Izod and Universal testers for metals	One year (complete calibration)	Frequent inspection by user.	BS EN 10045-2 Include verification using standard test pieces appropriate to required operating range(s).
	b) Charpy and Izod testers for plastics	One year (partial calibration) Five years (complete calibration)	Frequent inspection by user.	
	c) Notching tools		Check regularly and whenever reground.	
15.	Length Measuring Devices			
	a) Linearly Variable Differential Transformers		Daily or whenever used	Check against length standard such as a micrometer setting bar.
	b) Micrometers (hand)			
	i) For measurement of diameters smaller than 2.5mm and thickness less than 1.3mm	*Five years (complete)		
	ii) For measurement of diameters down to 2.5mm and thickness down to 1.3mm	*Five years (reference)		
	c) Rules	*Five years (reference)		
	d) Calipers – Vernier/Dial			
	i) Reference	*Three years (reference)		
ii) Working	*Annual		Against a reference length standard such as gauge bars.	
16.	Masses			
	a) Reference masses of integral construction stainless steel or nickel-chromium alloy	Five years		
	b) Masses of screw knob or sealed plug construction, made of stainless steel, nichrome, plated brass or other non-corrodible highly finished material	Three years		
c) Masses of cast iron, carbon steel, or unplated brass	*One year *Five years		if calibrated to 1 in 10^4 if calibrated to 1 in 10^3	

No.	Item of equipment	Calibration interval	Checking interval	General comments
17.	Nuclear Densometers	Two yearly	*Daily	Standard count (comparison against rolling average).
			*Six monthly	Drift and stability checks.
18.	Orifice Plates	Initial		PD ISO/TR 15377
			*Six months	Visual inspection for damage wear or contamination.
		Ten years		For orifice plates being used in window testing, a full recalibration is required after ten years.
19.	Ovens			
	a) Drying	*Five years	*Daily	For laboratories drying soils, a daily record of oven temperature is required. For laboratories drying aggregates, records showing temperature stability are required.
	b) Ageing	*Five years or less depending on permissible tolerances (temperature variations, recovery time, rate of ventilation)	Both drying and ageing ovens require full recalibration after major servicing.	
20.	Pressure Gauge Testers			
	a) Dead weight	Five years		
	b) Manometers i) liquid in glass ii) digital	Five years One year		
21.	Pressure and Vacuum Gauges			
	a) Test gauges for calibration of working gauges	One year		
	b) Working gauges subject to shock loading	*Six months or less depending on use		
	c) Working gauges not subject to shock loading	*One year		
22.	Manometers			
	a) Reference	Five years		Check against reference. Check fluid every three years.
	b) Working	*Three years		
	c) Digital	*One Year		

No.	Item of equipment	Calibration interval	Checking interval	General comments	
Proving devices for calibration of force testing machines					
<u>TYPE 1 – Elastic devices</u>					
23.	a) Dial gauge for deflection measurement	Two years			
	b) Micrometer screw for deflection measurement (mechanical or optical indication)	Five years			
	c) Electrical deflection measurement	Two years			
	<u>TYPE 2 – Proving levers</u>				
			Two years		
	<u>TYPE 3 – Weights</u>				
		Five years			
Sieves					
24.	a) Reference	* Initial			
	b) Working	*One year or less dependent on usage			
Soil Testing Machines					
25.	a) Force measurement	Two years			
	b) Displacement measurement	As for appropriate instrument (e.g. dial gauge, micrometer, LVDT)			
	c) Pressure measurement	As for pressure and vacuum gauges (hardness of rubber base)			
26.	Thickness Gauges (for compressible materials)	Two years		Dial gauge, dimensions and pressure of foot	
Squares					
27.	a) Reference	Five years		Against a reference square	
	b) Working	*Annual			
Stop Watches and Clocks					
28.	a) Electric	*Twelve months			
	b) Mechanical	*Three months			
Straight Edges					
29.	a) Reference	Five years			
	b) Strain rate meters	*Six months		using stop watch	
30.	a) Tachometer calibrators (Tuning devices)	Five years			
	b) Tachometers	One year			

No.	Item of equipment	Calibration interval	Checking interval	General comments
31.	Thermometers			
	a) Reference liquid-in-glass	Five years (complete)	*Six months	Check ice point immediately after initial calibration then at least every six months
	b) Working liquid-in-glass or alternatively	Five years (complete)	*Six months	Check ice point immediately after initial calibration then at least every six months Inter-compare with reference thermometer at points in the working range every six months
	c) Electronic (sensors that are thermocouples, thermistors, or other integrated circuit devices)	One year (full calibration)		
	d) Resistance	Five years (full calibration), or when ice point drift is more than five times the uncertainty of calibration.	Six months	Check at ice point before use or at least every six months. Working hand-held resistance thermometers can be checked using the alternative procedure above for glass thermometers.
32.	Volumetric glassware			
	a) Flasks, pipette, burettes and measuring cylinders used for reference purposes	*Five years		
	b) Working flasks, pipettes burettes, measuring cylinders	*On commissioning		Cross check by weighing with distilled water
	c) Density bottles	*Two years		

Table 2 b): Calibrations interval for reference standards and measuring instruments in non-destructive testing (NDT)

No.	Item of equipment	Calibration interval	Checking interval	General comments
A. Ultrasonic Testing				
1.	Probe and sensory electronics (setting up the assembly)		Each time before use	Ultrasonic standard calibration blocks
2.	Standard calibration blocks (material properties)		Initial	As per specific standard method (e.g. EN 27963, AWS, ASME or equivalent).
3.	Standard calibration blocks (surface conditions)		Each time before use	Visual examination for deterioration such as corrosion or mechanical damage.
4.	Reference standard calibration blocks (radius and other dimensional checks)	Every 5 years		By an accredited calibration laboratory or National Metrology Institute (NMI).
5.	Working standard calibration blocks (radius and other dimensional checks)		Intermediate checks every 2 years	By comparison with reference standard calibration block. If no reference standard calibration block is available, then by an accredited calibration laboratory or National Metrology Institute (NMI).
6.	Ultrasonic test sets <ul style="list-style-type: none"> • linearity of time base • linearity of equipment gains • sensitivity and signal to noise ratio • pulse duration. 		Verified weekly or each time the equipment is used	
7.	Ultrasonic probes and systems <ul style="list-style-type: none"> • probe index • probe beam angle • visual checks for damage 		The performance characteristics checked at least once per day or before use	
8.	Ultrasonic flaw detectors <ul style="list-style-type: none"> • linearity of time base • linearity of amplifier • accuracy of calibrated attenuator 		Verified at intervals not exceeding twelve months	

No.	Item of equipment	Calibration interval	Checking interval	General comments
B. Magnetic Particle Testing				
9.	Magnetic particle solution (visible/fluorescence)		Each batch	Valid manufacturer's certificate with conformance to a standard (e.g. BS, ASTM or EN).
10.	Magnetic inks (for aerosols)		Each batch	Valid manufacturer's certificate with conformance to a relevant standard. Flux indicators should be used to demonstrate the direction of flux.
11.	Magnetic particle concentration check		Each shift	As per specific standard method (e.g. ASTM, BS)
12.	Visible light level intensity checks on the test surface	As per specific standard method (e.g. ASTM, BS)		Check the level of illumination using a calibrated light meter each time before use.
13.	Black light level intensity check on the test surface			
14.	UV(A) light meter (Reference)	Yearly		By an accredited calibration laboratory or by National Metrology Institute (NMI)
15.	White light meter (Reference)			
16.	Permanent magnet and magnetic yokes		6 monthly	Check by measuring the lifting power or pull-off force in accordance with a relevant standard.
17.	Reference Weights (for checking strength of magnet)		Initial	Once calibrated for life. Calibrate by means of a calibrated balance.
18.	Gauss meter	Yearly		By an accredited calibration laboratory or National Metrology Institute (NMI).
19.	Ammeter			
C. Radiographic Testing				
20.	Gamma Ray – Source Size X-Ray – Focal Spot Size		Initial	Manufacturer's certification with official record of dimensions.
21.	Densitometer		90 days	Calibrate against a reference density strip, which is calibrated by an accredited calibration laboratory or National Metrology Institute (NMI).
22.	Film density strip	As manufacturer's recommendation, whichever is earlier.		By an accredited calibration laboratory or National Metrology Institute (NMI) where available. Note: Date of first usage of strip to be recorded.
23.	Survey meters	Yearly		As required by the Atomic Energy Licensing Board (AELB).
24.	Gamma projector	Yearly		As required by the Atomic Energy Licensing Board (AELB).
25.	X-ray machine	Yearly		As required by the Atomic Energy Licensing Board (AELB).
26.	Digital Radiographic. * Optical Density Step Wedges * Optical Line Pair Test Pattern	Once every 5 years		By an accredited calibration laboratory or National Metrology Institute (NMI).

No.	Item of equipment	Calibration interval	Checking interval	General comments
D. Penetrant Testing				
27.	Non-fluorescent (aerosol) penetrant dyes		Each batch	Valid manufacturer's certificate with conformance to a relevant standard. Where possible verification against Penetrant Comparator Block.
28.	Fluorescent dyes		Each batch / per work day	Manufacturer's certificate with conformance to a relevant standard. Where possible verification against Penetrant Comparator Block.
29.	Reference UV(A) light meter	Yearly		By an accredited calibration laboratory or 'National Metrology Institute (NMI)', where available.
30.	Reference White light meter			
31.	UV(A) light meter (Working)		90 days	By comparison with reference light meter
32.	White light meter (Working)			
E. Eddy current Testing				
33.	Reference Specimen	Initial		Manufacturer's certification, customer's requirements or measurement certificate.
34.	Equipment		Before usage and subsequently after 8 hours of usage	Verified against a reference specimen

Bibliography

1. MS ISO/IEC 17025 - General requirements for the competence of testing and calibration laboratories.
2. ISO/IEC 17043 - Conformity assessment - General requirements for proficiency testing.
3. ISO 9712 - Non-destructive testing - Qualification and certification of NDT personnel.
4. ISO 10012 - Measurement management systems - Requirements for measurement processes and measuring equipment.
5. JCGM 100 - Evaluation of measurement data - Guide to the expression of uncertainty in measurement.
6. ILAC-G24 - Guidelines for the determination of calibration intervals of measuring instruments.
7. EA Document - EA-4/15 G Accreditation for Non-Destructive Testing.
8. IANZ Specific Criteria for Accreditation - Mechanical Testing.
9. IANZ Supplementary Criteria for Accreditation - Mechanical Testing Laboratories (Non- Destructive Testing).
10. SAC-SINGLAS - NDT 003 - Quality Assurance of Equipment Commonly Used in Non-Destructive Testing laboratories.

ACKNOWLEDGEMENTS

1. Dr. Afidah Abu Bakar (Chairman) Standards Malaysia
 2. Ms. Rosnieh Eggat (Secretariat) Standards Malaysia
 3. Dr. Mohamad Pauzi bin Ismail Standards Malaysia
 4. Mr. Wong Siew Kwan Standards Malaysia
 5. Dr. Norhayati Moris Standards Malaysia
 6. Mr. Mohamad Akhir bin Abdullah Ikram QA Services Sdn. Bhd.
 7. Ms. Seri Banun Sujangi MMC Gamuda
 8. Mr. Mohd Zahari Bin Shariffudin Makmal Ujian Bahan Binaan,
CIDB Holdings Sdn. Bhd.
 9. Mr. Syuhaizat Md. Ali Lembaga Getah Malaysia (LGM)
 10. Mr. Dalha Rahmat SIRIM QAS International Sdn Bhd.
-